

SCHEDULE 2

METHODS OF ANALYSIS

3b.

DETERMINATION OF NITRIC AND AMMONIACAL NITROGEN — ARND METHOD

Control test

7.4 Before analysis, check that apparatus is working properly and that the correct technique is applied using a freshly prepared solution of sodium nitrate (4.13) containing 0.050 g to 0.150 g nitrogen depending on the variant chosen.