

SCHEDULE 1

ACCURACY CLASSIFICATION OF NON-AUTOMATIC WEIGHING MACHINES

PART I

GENERAL

9. For machines fitted with a graduated verification device the scale interval of such an incorporated device must be at most equal to one-fifth of the scale interval of the machine.