

SCHEDULE 4

METHODS OF ANALYSIS

PART B

Chemical and indicator parameters

1.—(1) Subject to paragraph 3, for a parameter in Table 1, the method of analysis used must, as a minimum, be capable of measuring concentrations equal to the parametric value for the parameter with—

- (a) a limit of quantification of 30 % or less of the parametric value; and
- (b) an uncertainty of measurement as specified in Table 1.

(2) The result of the analysis for a parameter in Table B or Table C must be expressed using at least the same number of significant figures as the parametric value for the parameter in the table.