
STATUTORY INSTRUMENTS

1963 No. 1710

The Weights and Measures Regulations 1963

PART III

LIQUID CAPACITY MEASURES

Materials and principles of construction

23. Liquid capacity measures made of glass, other than apothecaries measures, shall have their maximum purported values defined either—

- (a) by the brim of the measure; or
- (b) by a line not less than 2 inches in length, and distant not less than half an inch nor more than 1½ inches from the brim.