

---

STATUTORY INSTRUMENTS

---

**1963 No. 1710**

The Weights and Measures Regulations 1963

PART IX

SPRING BALANCES

*Testing*

**95.** Spring balances may be tested for efficiency or ability to recover by leaving on them for a period not exceeding 24 hours, a load equal to the capacity of the instrument; and then, after the expiration of a further period of 4 hours, by testing for accuracy.