

Status: This is the original version (as it was originally made). This item of legislation is currently only available in its original format. The electronic version of this UK Statutory Instrument has been contributed by Westlaw and is taken from the printed publication. **Read more**

SCHEDULE 2

PART I

PRESCRIBED LIMITS OF ERROR ON THE TESTING OF WEIGHING INSTRUMENTS

1. Subject to paragraphs 2 and 3 of this Part of this Schedule, the prescribed limits of error for weighing instruments shall be those specified in Parts II to XI of this Schedule:

Provided that in the case of any weighing instrument of a capacity not so specified, the prescribed limits of error shall be the amounts proportionate to those so specified for an instrument of the same type, class or description.

2. In the case of any weighing instrument which weighs in units of the metric system and for which no limits of error are specified in terms of those units, the prescribed limits of error shall be the amounts in terms of metric units equivalent to those specified in terms of imperial units in the relevant Part of this Schedule with respect to an instrument of the same capacity, type, class or description.

3. In the case of any weighing instrument of the self-indicating or semi-self-indicating type, the prescribed limit of error, in excess or in deficiency, shall be either—

- (a) the appropriate amount specified in the relevant Part of this Schedule for the instrument concerned, or
- (b) the amount corresponding to one half of the smallest interval between consecutive graduations on the scale or dial of the instrument,

whichever is the less.

PART II

BEAM SCALES AND BALANCES

1. *Beam scales marked "Class B".*

<i>Capacity of instrument Upon passing as fit for use for trade</i>	<i>Weight to be added to test sensitiveness when fully loaded</i>		<i>Error in excess or in deficiency when fully loaded</i>	
	<i>In relation to the obliteration of stamps</i>	<i>Upon passing as fit for use for trade</i>	<i>In relation to the obliteration of stamps</i>	
1 ounce	#grain	1#grain	# grain	1# grain
8 ounces	1 grain	3 grains	1 grain	2 grains
1 pound	1 grain	3 grains	1 grains	2 grains
2 pounds	1½ grains	4½ grains	2 grains	4 grains
4 pounds	3 grains	9 grains	4 grains	8 grains
7 pounds	4 grains	12 grains	6 grains	12 grains
10 pounds	6 grains	18 grains	9 grains	18 grains

Status: This is the original version (as it was originally made). This item of legislation is currently only available in its original format. The electronic version of this UK Statutory Instrument has been contributed by Westlaw and is taken from the printed publication. **Read more**

<i>Capacity of instrument Upon passing as fit for use for trade</i>	<i>Weight to be added to test sensitiveness when fully loaded</i>		<i>Error in excess or in deficiency when fully loaded</i>	
	<i>In relation to the obliteration of stamps</i>	<i>Upon passing as fit for use for trade</i>	<i>In relation to the obliteration of stamps</i>	
14 pounds	8 grains	24 grains	12 grains	24 grains
28 pounds	15 grains	45 grains	22 grains	44 grains
56 pounds	25 grains	75 grains	40 grains	80 grains
112 pounds	1½ drams	4½ drams	2½ drams	5 drams
224 pounds	2½ drams	7½ drams	3½ drams	7 drams
Above 2 hundredweight	Add ½ dram for each hundred- weight of capacity	Add 1½ drams for each hundred- weight of capacity	Add 1 dram for each hundred- weight of capacity	Add 2 drams for each hundred- weight of capacity

2. *Beam scales marked “Class C”.*

<i>Capacity of instrument Upon passing as fit for use for trade</i>	<i>Weight to be added to test sensitiveness when fully loaded</i>		<i>Error in excess or in deficiency when fully loaded</i>	
	<i>In relation to the obliteration of stamps</i>	<i>Upon passing as fit for use for trade</i>	<i>In relation to the obliteration of stamps</i>	
1 ounce	# grain	1# grain	# grain	1# grains
8 ounces	3 grains	6 grains	3 grains	6 grains
1 pound	3 grains	6 grains	3 grains	6 grains
2 pounds	4½ grains	9 grains	6 grains	12 grains
4 pounds	9 grains	18 grains	12 grains	24 grains
7 pounds	12 grains	24 grains	18 grains	36 grains
10 pounds	18 grains	36 grains	27 grains	54 grains
14 pounds	24 grains	48 grains	36 grains	72 grains
28 pounds	45 grains	90 grains	66 grains	132 grains
56 pounds	75 grains	150 grains	120 grains	240 grains
112 pounds	4½ drams	9 drams	7½ drams	15 drams
224 pounds	7½ drams	15 drams	10½ drams	21 drams
Above 2 hundredweight	Add 1½ drams for each hundred-	Add 3 drams for each hundred-	Add 3 drams for each hundred-	Add 6 drams for each hundred-

Status: This is the original version (as it was originally made). This item of legislation is currently only available in its original format. The electronic version of this UK Statutory Instrument has been contributed by Westlaw and is taken from the printed publication. **Read more**

<i>Capacity of instrument Upon passing as fit for use for trade</i>	<i>Weight to be added to test sensitiveness when fully loaded</i>		<i>Error in excess or in deficiency when fully loaded</i>	
	<i>In relation to the obliteration of stamps weight of capacity</i>	<i>Upon passing as fit for use for trade weight of capacity</i>	<i>In relation to the obliteration of stamps weight of capacity</i>	<i>weight of capacity</i>

3. Balances.

<i>Capacity of instrument Upon passing as fit for use for trade</i>	<i>Weight to be added to test sensitiveness when fully loaded</i>		<i>Error in excess or in deficiency when fully loaded</i>	
	<i>In relation to the obliteration of stamps</i>	<i>Upon passing as fit for use for trade</i>	<i>In relation to the obliteration of stamps</i>	
1 ounce	120 grain	320 grain	110 grain	# grain
1 pound	110 grain	310 grain	# grain	# grain
7 pounds	½ grain	1½ grains	1 grain	2 grains
56 pounds	1½ grains	4½ grains	2 grains	4 grains

PART III

COUNTER MACHINES

<i>Capacity of machine Upon passing as fit for use for trade</i>	<i>Weight to be added to test sensitiveness when fully loaded</i>		<i>Error in excess or in deficiency when fully loaded</i>	
	<i>In relation to the obliteration of stamps</i>	<i>Upon passing as fit for use for trade</i>	<i>In relation to the obliteration of stamps</i>	
1 pound	20 grains	60 grains	30 grains	60 grains
2 pounds	28 grains	84 grains	1½ drams	3 drams
4 pounds	40 grains	120 grains	2 drams	4 drams
7 pounds	2 drams	6 drams	3 drams	6 drams
10 pounds	2½ drams	7½ drams	3½ drams	7 drams
14 pounds	3 drams	9 drams	4½ drams	9 drams
28 pounds	4 drams	12 drams	6 drams	12 drams
56 pounds	6 drams	18 drams	9 drams	18 drams
1 hundredweight	8 drams	24 drams	16 drams	32 drams

Status: This is the original version (as it was originally made). This item of legislation is currently only available in its original format. The electronic version of this UK Statutory Instrument has been contributed by Westlaw and is taken from the printed publication. **Read more**

PART IV
SPRING BALANCES

<i>Capacity of spring balance Upon passing as fit for use for trade</i>	<i>Error in excess or in deficiency when fully loaded</i>	
	<i>In relation to the obliteration of stamps</i>	
1 pound	30 grains	60 grains
2 pounds	1½ drams	3 drams
3 pounds	1½ drams	3 drams
4 pounds	2 drams	4 drams
5 pounds	2½ drams	5 drams
6 pounds	2½ drams	5 drams
7 pounds	3 drams	6 drams
10 pounds	3½ drams	7 drams
11 pounds	3½ drams	7 drams
12 pounds	4 drams	8 drams
13 pounds	4 drams	8 drams
14 pounds	4½ drams	9 drams
15 pounds	4½ drams	9 drams
20 pounds	5 drams	10 drams
21 pounds	5 drams	10 drams
22 pounds	5 drams	10 drams
23 pounds	5½ drams	11 drams
24 pounds	5½ drams	11 drams
25 pounds	5½ drams	11 drams
26 pounds	5½ drams	11 drams
27 pounds	6 drams	12 drams
28 pounds	6 drams	12 drams
29 pounds	6 drams	12 drams
30 pounds	6 drams	12 drams
40 pounds and above	The weight corresponding to ¼ of the interval between consecutive graduations	The weight corresponding to ½ of the interval between consecutive graduations

Status: This is the original version (as it was originally made). This item of legislation is currently only available in its original format. The electronic version of this UK Statutory Instrument has been contributed by Westlaw and is taken from the printed publication. **Read more**

PART V STEELYARDS

<i>Weight to be added to test sensitiveness when fully loaded</i>		<i>Error in excess or in deficiency when fully loaded</i>		
<i>Capacity of steelyard</i>	<i>Upon passing as fit for use for trade</i>	<i>In relation to the obliteration of stamps</i>	<i>Upon passing as fit for use for trade</i>	<i>In relation to the obliteration of stamps</i>
56 pounds	12 drams	36 drams	18 drams	36 drams
1 hundredweight	1 ounce	3 ounces	2 ounces	4 ounces
3 hundredweight	2 ounces	6 ounces	4 ounces	8 ounces
5 hundredweight	3 ounces	9 ounces	6 ounces	12 ounces
7 hundredweight	4 ounces	12 ounces	8 ounces	16 ounces
10 hundredweight	6 ounces	18 ounces	12 ounces	24 ounces
20 hundredweight	10 ounces	30 ounces	20 ounces	40 ounces
30 hundredweights	13 ounces	39 ounces	26 ounces	52 ounces
40 hundredweights	16 ounces	48 ounces	32 ounces	64 ounces
50 hundredweights	20 ounces	60 ounces	40 ounces	80 ounces

PART VI DEAD-WEIGHT MACHINES

<i>Capacity of machine</i>	<i>Vibrating weighing instruments</i>		<i>Accelerating weighing instruments</i>				<i>Weight required to bring back the beam from position of greatest displacement when fully loaded</i>
	<i>Weight to be added to test sensitiveness when fully loaded</i>		<i>Error in excess or in deficiency when fully loaded</i>		<i>Error in excess or in deficiency when fully loaded</i>		
<i>Upon passing as fit for the for trade</i>	<i>In relation to the obliteration of stamps</i>	<i>Upon passing as fit for trade</i>	<i>In relation to the obliteration of stamps</i>	<i>Upon passing as fit for trade</i>	<i>In relation to the obliteration of stamps</i>	<i>Upon passing as fit for trade</i>	
1 hundredweight	½ ounce	1½ ounces	1 ounce	2 ounces	1 ounce	2 ounces	2 ounces

Status: This is the original version (as it was originally made). This item of legislation is currently only available in its original format. The electronic version of this UK Statutory Instrument has been contributed by Westlaw and is taken from the printed publication. **Read more**

<i>Capacity of machine Upon passing as fit for the for trade</i>	<i>Vibrating weighing instruments</i>		<i>Accelerating weighing instruments</i>				<i>Weight required to bring back the beam from position of greatest displacement when fully loaded</i>
	<i>Weight to be added to test sensitiveness when fully loaded</i>		<i>Error in excess or in deficiency when fully loaded</i>		<i>Error in excess or in deficiency when fully loaded</i>		
	<i>In relation to the obliteration of stamps</i>	<i>Upon passing as fit for use for trade</i>	<i>In relation to the obliteration of stamps</i>	<i>Upon passing as fit for use for trade</i>	<i>In relation to the obliteration of stamps</i>	<i>Upon passing as fit for use for trade</i>	
3 hundredweight	1 ounce	3 ounces	2 ounces	4 ounces	2 ounces	4 ounces	4 ounces
5 hundredweight	1½ ounces	4½ ounces	3 ounces	6 ounces	3 ounces	6 ounces	6 ounces
7 hundredweight	2 ounces	6 ounces	4 ounces	8 ounces	4 ounces	8 ounces	8 ounces
10 hundredweight	3 ounces	9 ounces	6 ounces	12 ounces	6 ounces	12 ounces	12 ounces
20 hundredweight	5 ounces	15 ounces	10 ounces	20 ounces	10 ounces	20 ounces	20 ounces
30 hundredweight	6½ ounces	19½ ounces	13 ounces	26 ounces	13 ounces	26 ounces	26 ounces
40 hundredweight	8 ounces	24 ounces	16 ounces	32 ounces	16 ounces	32 ounces	32 ounces
50 hundredweight	10 ounces	30 ounces	20 ounces	40 ounces	20 ounces	40 ounces	40 ounces

Status: This is the original version (as it was originally made). This item of legislation is currently only available in its original format. The electronic version of this UK Statutory Instrument has been contributed by Westlaw and is taken from the printed publication. **Read more**

PART VII

PLATFORM WEIGHING MACHINES

<i>Vibrating weighing instruments</i>		<i>Accelerating weighing instruments</i>				<i>Machines with dials</i>			
<i>Capacity of machine</i>	<i>Weight to be added to test sensitiveness when fully loaded</i>	<i>Error in excess or in deficiency when fully loaded</i>	<i>Error in excess or in deficiency when fully loaded</i>	<i>Error in excess or in deficiency when fully loaded</i>	<i>Error in excess or in deficiency when fully loaded</i>	<i>Weight required to bring back the steelyard indicator from position of greatest displacement when fully loaded must not exceed—</i>	<i>Error in excess or in deficiency when fully loaded</i>	<i>Error in excess or in deficiency when fully loaded</i>	<i>Error in excess or in deficiency when fully loaded</i>
<i>Upon passing as fit for trade</i>	<i>In relation to the obliteration of stamps</i>	<i>Upon passing as fit for trade</i>	<i>In relation to the obliteration of stamps</i>	<i>Upon passing as fit for trade</i>	<i>In relation to the obliteration of stamps</i>	<i>Upon passing as fit for trade</i>	<i>Upon passing as fit for trade</i>	<i>In relation to the obliteration of stamps</i>	<i>In relation to the obliteration of stamps</i>
1 hundred-weight	½ ounce	1½ ounces	1 ounce	2 ounces	1 ounce	2 ounces	2 ounces	2 ounces	4 ounces
3 hundred-weight	1 ounce	3 ounces	2 ounces	4 ounces	2 ounces	4 ounces	4 ounces	4 ounces	8 ounces
5 hundred-weight	1½ ounces	4½ ounces	3 ounces	6 ounces	3 ounces	6 ounces	6 ounces	6 ounces	12 ounces
7 hundred-weight	2 ounces	6 ounces	4 ounces	8 ounces	4 ounces	8 ounces	8 ounces	8 ounces	16 ounces
10 hundred-weight	3 ounces	9 ounces	6 ounces	12 ounces	6 ounces	12 ounces	12 ounces	12 ounces	24 ounces
20 hundred-weight	5 ounces	15 ounces	10 ounces	20 ounces	10 ounces	20 ounces	20 ounces	20 ounces	40 ounces
30 hundred-weight	6½ ounces	19½ ounces	13 ounces	26 ounces	13 ounces	26 ounces	26 ounces	26 ounces	52 ounces

Status: This is the original version (as it was originally made). This item of legislation is currently only available in its original format. The electronic version of this UK Statutory Instrument has been contributed by Westlaw and is taken from the printed publication. **Read more**

<i>Vibrating weighing instruments</i>		<i>Accelerating weighing instruments</i>		<i>Machines with dials</i>	
<i>Capacity of machine</i>	<i>Weight to be added to test sensitiveness when fully loaded</i>	<i>Error in excess or in deficiency when fully loaded</i>	<i>Error in excess or in deficiency when fully loaded</i>	<i>Weight required to bring back the steelyard indicator from position of greatest displacement when fully loaded must not exceed—</i>	<i>Error in excess or in deficiency when fully loaded</i>
<i>Upon passing as fit for trade</i>	<i>In relation to the obliteration of stamps</i>	<i>Upon passing as fit for trade</i>	<i>In relation to the obliteration of stamps</i>	<i>Upon passing as fit for trade</i>	<i>In relation to the obliteration of stamps</i>
40	8	24	16	32	16
hundred-weight	ounces	ounces	ounces	ounces	ounces
50	10	30	20	40	20
hundred-weight	ounces	ounces	ounces	ounces	ounces

PART VIII

SELF-INDICATING PIT-BANK WEIGHING MACHINES

<i>Capacity of machine</i>	<i>Error in excess or in deficiency when fully loaded</i>	
<i>Upon passing as fit for use for trade</i>	<i>In relation to the obliteration of stamps</i>	
1 hundred weight	6 ounces	12 ounces
2 hundred weight	9 ounces	1 pound 2 ounces
3 hundred weight	12 ounces	1 pound 8 ounces
4 hundred weight	15 ounces	1 pound 14 ounces
5 hundred weight	1 pound 2 ounces	2 pounds 4 ounces
7 hundred weight	1 pound 8 ounces	3 pounds 0 ounces
10 hundred weight	2 pounds 4 ounces	4 pounds 8 ounces

Status: This is the original version (as it was originally made). This item of legislation is currently only available in its original format. The electronic version of this UK Statutory Instrument has been contributed by Westlaw and is taken from the printed publication. **Read more**

<i>Capacity of machine</i>	<i>Error in excess or in deficiency when fully loaded</i>	
<i>Upon passing as fit for use for trade</i>	<i>In relation to the obliteration of stamps</i>	
12 hundred weight	2 pounds 9 ounces	5 pounds 2 ounces
15 hundred weight	3 pounds 0 ounces	6 pounds 0 ounces
20 hundred weight	3 pounds 12 ounces	7 pounds 8 ounces
30 hundred weight	4 pounds 14 ounces	9 pounds 12 ounces
40 hundred weight	6 pounds 0 ounces	12 pounds 0 ounces
50 hundred weight	7 pounds 8 ounces	15 pounds 0 ounces

PART IX WEIGHBRIDGES

<i>Capacity of machine</i>	<i>Vibrating weighing instruments without dials</i>		<i>Accelerating weighing instruments without dials</i>		<i>Instruments with dials</i>				
	<i>Weight to be added to test sensitiveness when fully loaded</i>	<i>Error in excess or in deficiency when fully loaded</i>	<i>Error in excess or in deficiency when fully loaded</i>	<i>Weight required to bring back the steelyard indicator from position of greatest displacement when fully loaded must not exceed—</i>	<i>Error in excess or in deficiency when fully loaded</i>				
<i>Upon passing as fit for use for trade</i>	<i>In relation to the obliteration of stamps</i>	<i>Upon passing as fit for trade</i>	<i>In relation to the obliteration of stamps</i>	<i>Upon passing as fit for trade</i>	<i>In relation to the obliteration of stamps</i>	<i>Upon passing as fit for trade</i>	<i>In relation to the obliteration of stamps</i>		
1 ton	1½ pounds	4½ pounds	1½ pounds	3 pounds	1½ pounds	3 pounds	4 pounds	3 pounds	6 pounds
2 tons	2 pounds	6 pounds	2 pounds	4 pounds	2 pounds	4 pounds	5 pounds	4 pounds	8 pounds
5 tons	3½ pounds	10 ½ pounds	4 pounds	8 pounds	4 pounds	8 pounds	10 pounds	8 pounds	16 pounds

Status: This is the original version (as it was originally made). This item of legislation is currently only available in its original format. The electronic version of this UK Statutory Instrument has been contributed by Westlaw and is taken from the printed publication. **Read more**

<i>Capacity of machine</i>	<i>Vibrating weighing instruments without dials</i>			<i>Accelerating weighing instruments without dials</i>			<i>Instruments with dials</i>		
	<i>Weight to be added to test sensitiveness when fully loaded</i>	<i>Error in excess or in deficiency when fully loaded</i>	<i>Error in excess or in deficiency when fully loaded</i>	<i>Weight required to bring back the steelyard indicator from position of greatest displacement when fully loaded must not exceed—</i>	<i>Error in excess or in deficiency when fully loaded</i>	<i>Weight required to bring back the steelyard indicator from position of greatest displacement when fully loaded must not exceed—</i>	<i>Error in excess or in deficiency when fully loaded</i>	<i>Weight required to bring back the steelyard indicator from position of greatest displacement when fully loaded must not exceed—</i>	<i>Error in excess or in deficiency when fully loaded</i>
	<i>Upon passing as fit for use for trade</i>	<i>In relation to the obliteration of stamps</i>	<i>Upon passing as fit for use for trade</i>	<i>In relation to the obliteration of stamps</i>	<i>Upon passing as fit for use for trade</i>	<i>In relation to the obliteration of stamps</i>	<i>Upon passing as fit for use for trade</i>	<i>In relation to the obliteration of stamps</i>	<i>Upon passing as fit for use for trade</i>
10 tons	5 pounds	15 pounds	6 pounds	12 pounds	6 pounds	12 pounds	15 pounds	12 pounds	24 pounds
20 tons	7 pounds	21 pounds	10 pounds	20 pounds	10 pounds	20 pounds	25 pounds	20 pounds	40 pounds
25 tons	8 pounds	24 pounds	12 pounds	24 pounds	12 pounds	24 pounds	30 pounds	24 pounds	48 pounds
30 tons	8½ pounds	25½ pounds	13½ pounds	27 pounds	13½ pounds	27 pounds	34 pounds	27 pounds	54 pounds
35 tons	9 pounds	27 pounds	15 pounds	30 pounds	15 pounds	30 pounds	37 pounds	30 pounds	60 pounds
40 tons	9½ pounds	28½ pounds	16 pounds	32 pounds	16 pounds	32 pounds	40 pounds	32 pounds	64 pounds
50 tons	10 pounds	30 pounds	18 pounds	36 pounds	18 pounds	36 pounds	45 pounds	36 pounds	52 pounds
75 tons	12 pounds	36 pounds	23 pounds	46 pounds	23 pounds	46 pounds	58 pounds	46 pounds	92 pounds
100 tons	14 pounds	42 pounds	28 pounds	56 pounds	28 pounds	56 pounds	70 pounds	56 pounds	112 pounds
200 tons	18 pounds	54 pounds	42 pounds	84 pounds	42 pounds	84 pounds	105 pounds	84 pounds	168 pounds

PART X

CRANE WEIGHING MACHINES

Crane weighing machines constructed upon other than the hydraulic principle

<i>Capacity of machine Upon passing as fit for use for trade</i>	<i>Machines with steelyard indicators</i>		<i>Machines with dials</i>			
	<i>Weight to be added to test sensitiveness when fully loaded</i>		<i>Error in excess or in deficiency when fully loaded</i>		<i>Error in excess or in deficiency when fully loaded</i>	
	<i>In relation to the obliteration of stamps</i>	<i>Upon passing as fit for use for trade</i>	<i>In relation to the obliteration of stamps</i>	<i>Upon passing as fit for use for trade</i>	<i>In relation to the obliteration of stamps</i>	<i>In relation to the obliteration of stamps</i>
1 hundredweight	½ ounce	1½ ounces	1 ounces	2 ounces	2 ounces	4 ounces
5 hundredweight	1½ ounces	4½ ounces	3 ounces	6 ounces	6 ounces	12 ounces
10 hundredweight	3 ounces	9 ounces	6 ounces	12 ounces	12 ounces	1½ pounds
1 ton	1½ pounds	4½ pounds	1½ pounds	3 pounds	3 pounds	6 pounds
2 tons	2 pounds	6 pounds	2 pounds	4 pounds	4 pounds	8 pounds
5 tons	3½ pounds	10½ pounds	4 pounds	8 pounds	8 pounds	16 pounds
10 tons	5 pounds	15 pounds	6 pounds	12 pounds	12 pounds	24 pounds
20 tons	7 pounds	21 pounds	10 pounds	20 pounds	20 pounds	40 pounds
25 tons	8 pounds	24 pounds	12 pounds	24 pounds	24 pounds	48 pounds
30 tons	8½ pounds	25½ pounds	13½ pounds	27 pounds	27 pounds	54 pounds
35 tons	9 pounds	27 pounds	15 pounds	30 pounds	30 pounds	60 pounds
40 tons	9½ pounds	28½ pounds	16 pounds	32 pounds	32 pounds	64 pounds
50 tons	10 pounds	30 pounds	18 pounds	36 pounds	36 pounds	72 pounds
75 tons	12 pounds	36 pounds	23 pounds	46 pounds	46 pounds	92 pounds
100 tons	14 pounds	42 pounds	28 pounds	56 pounds	56 pounds	112 pounds
200 tons	18 pounds	54 pounds	42 pounds	84 pounds	84 pounds	168 pounds

Crane weighing machines constructed upon the hydraulic principle

2. An amount equal to one-half of the weight represented by the interval between consecutive graduations.

Status: This is the original version (as it was originally made). This item of legislation is currently only available in its original format. The electronic version of this UK Statutory Instrument has been contributed by Westlaw and is taken from the printed publication. **Read more**

PART XI
AUTOMATIC WEIGHING MACHINES

<i>Error in relation to the passing as fit for use for trade</i>		<i>Error in relation to the obliteration of stamps</i>					
<i>For the purposes of Test A</i>		<i>For the purposes of Test A</i>		<i>For the purposes of Test B</i>		<i>For the purposes of Test B</i>	
<i>Description of machine</i>	<i>Capacity of machine</i>	<i>When tested by means of test loads</i>	<i>When tested by means of appropriate weights</i>	<i>For the purposes of Test B</i>	<i>When tested by means of test loads</i>	<i>When tested by means of appropriate weights</i>	<i>For the purposes of Test B</i>
<i>Column 1</i>	<i>Column 2</i>	<i>Column 3</i>	<i>Column 4</i>	<i>Column 5</i>	<i>Column 6</i>	<i>Column 7</i>	<i>Column 8</i>
All automatic weighing machines, other than those hereinafter described in columns 1 and 2 of this table	Under 10 pounds	Under ½ per cent., in excess only, of the purported weight of each test load	Prescribed limit of error applicable to the type, class or description of the weighing instrument to which the machine most nearly relates	½ per cent., in excess only, of the purported weight of the test load	1 per cent., in excess only, of the purported weight of the test load	Prescribed limit of error applicable to the type, class or description of the weighing instrument to which the machine most nearly relates	1 per cent., in excess only, of the purported weight of the test load
10 pounds or more	½ per cent., in excess only, of the purported weight of each test load	Prescribed limit of error applicable to the type, class or description of the weighing instrument to which the machine most nearly relates	Test not applicable	1 per cent., in excess or in deficiency, of the purported weight of each test load	Prescribed limit of error applicable to the type, class or description of the weighing instrument to which the machine most nearly relates	Test not applicable	Test not applicable

Status: This is the original version (as it was originally made). This item of legislation is currently only available in its original format. The electronic version of this UK Statutory Instrument has been contributed by Westlaw and is taken from the printed publication. **Read more**

<i>Error in relation to the passing as fit for use for trade</i>		<i>Error in relation to the obliteration of stamps</i>					
<i>For the purposes of Test A</i>		<i>For the purposes of Test A</i>		<i>For the purposes of Test B</i>		<i>For the purposes of Test B</i>	
<i>Description of machine</i>	<i>Capacity of machine</i>	<i>When tested by means of test loads</i>	<i>When tested by means of appropriate weights</i>	<i>When tested by means of the direct application of appropriate weights</i>	<i>When tested by means of test loads</i>	<i>When tested by means of appropriate weights</i>	<i>When tested by means of the direct application of appropriate weights</i>
<i>Column 1</i>	<i>Column 2</i>	<i>Column 3</i>	<i>Column 4</i>	<i>Column 5</i>	<i>Column 6</i>	<i>Column 7</i>	<i>Column 8</i>
Automatic weighing machines for use only for weighing grain	10 pounds or more	¼ per cent., in excess or deficiency, of the purported weight of each test load	Prescribed limit of error applicable to the type, class or description of the weighing instrument to which the machine most nearly relates	Test not applicable	½ per cent., in excess or deficiency, of the purported weight of each test load	Prescribed limit of error applicable to the type, class or description of weighing instrument to which the machine most nearly relates	Test not applicable
Automatic weighing machines for use only for weighing solid fuel	2 hundredweight or less	2 per cent., in excess or deficiency, of the purported weight of each test load	Prescribed limit of error applicable to the type, class or description of the weighing instrument to which the machine most nearly relates	Test not applicable	2 per cent., in excess only, of the purported weight of each test load	Prescribed limit of error applicable to the type, class or description of the weighing instrument to which the machine most nearly relates	Test not applicable
Automatic weighing machines for use	Any capacity	20 per cent., in excess only, of	Prescribed limit of error applicable	Test not applicable	20 per cent., in excess only, of	Prescribed limit of error applicable	Test not applicable

Status: This is the original version (as it was originally made). This item of legislation is currently only available in its original format. The electronic version of this UK Statutory Instrument has been contributed by Westlaw and is taken from the printed publication. **Read more**

<i>Error in relation to the passing as fit for use for trade</i>		<i>Error in relation to the obliteration of stamps</i>					
<i>For the purposes of Test A</i>		<i>For the purposes of Test A</i>		<i>For the purposes of Test B</i>		<i>For the purposes of Test B</i>	
<i>Description of machine</i>	<i>Capacity of machine</i>	<i>When tested by means of test loads</i>	<i>When tested by means of the direct application of appropriate weights</i>	<i>For the purposes of Test B</i>	<i>When tested by means of test loads</i>	<i>When tested by means of the direct application of appropriate weights</i>	<i>For the purposes of Test B</i>
<i>Column 1</i>	<i>Column 2</i>	<i>Column 3</i>	<i>Column 4</i>	<i>Column 5</i>	<i>Column 6</i>	<i>Column 7</i>	<i>Column 8</i>
only for weighing potato crisps		the total purported weight of 20 test loads	to the type, class or description of the weighing instrument to which the machine most nearly relates		the total purported weight of 20 test loads	to the type, class or description of weighing instrument to which the machine most nearly relates	
Totalising weighing machines,	Any capacity	½ per cent., in excess or in deficiency, of the total purported weight of 40 test loads	Prescribed limit of error applicable to the type, class or description of the weighing instrument to which the machine most nearly relates	Test not applicable	1 per cent., in excess or in deficiency, of the total purported weight of 40 test loads	Prescribed limit of error applicable to the type, class or description of weighing instrument to which the machine most nearly relates	Test not applicable