
STATUTORY INSTRUMENTS

1996 No. 797

The Weighing Equipment (Filling and Discontinuous Totalising Automatic Weighing Machines) (Amendment) Regulations 1996

Amendment

13. The principal Regulations are hereby further amended in Schedule 3, by the substitution for Table 2 of the following Table—

“DISCONTINUOUS
TOTALISERS

Table 2

Class of accuracy of discontinuous totaliser	Limits of error relative to weight of material used in tests	
	In relation to passing as fit for use for trade	In relation to obliteration of the stamp
Class 0.2	Plus or minus 0.1%	Plus or minus 0.2%
Class 0.5	Plus or minus 0.25%	Plus or minus 0.5%
Class 1	Plus or minus 0.5%	Plus or minus 1%
Class 2	Plus or minus 1%	Plus or minus 2%”