

Status: This is the original version (as it was originally made). This item of legislation is currently only available in its original format.

SCHEDULE 3

(Annex III to Directive 97/68/EC)
TEST PROCEDURE

Status: This is the original version (as it was originally made). This item of legislation is currently only available in its original format.

Appendix 1

Repeatability

1.4.1.2. The repeatability, defined as 2.5 times the standard deviation of 10 repetitive responses to a given calibration or span gas, must be no greater than $\pm 1\%$ of full scale concentration for each range used above 155 ppm (or ppm C) or $\pm 2\%$ of each range used below 155 ppm (or ppm C).