
STATUTORY INSTRUMENTS

2000 No. 3236

The Non-automatic Weighing Instruments Regulations 2000

PART III

USE FOR TRADE OF NON-AUTOMATIC WEIGHING INSTRUMENTS

Instruments marked with temperature range

30. Where an instrument is marked with a temperature range, a person shall not use the instrument for trade at temperatures outside that range.