

SCHEDULE 1

ACCURACY CLASSIFICATION OF NON-AUTOMATIC WEIGHING MACHINES

PART V

MACHINES DESIGNATED CLASS III

16. A machine which is—

- (a) made in accordance with an approved pattern and which is or could be marked **III**, or
- (b) first stamped before 1st November 1988 in accordance with the provisions for crane weighing machines constructed upon the hydraulic principle in the Weights and Measures Regulations 1963, or
- (c) of the type referred to as “approximate weighers” and first stamped before 1st November 1988, or
- (d) of a type referred to as “common form” and which was first stamped before 1st November 1988,

is a Class III machine.

17. The specifications for non-graduated and graduated Class III machines are given in Table 4.

Table 4

<i>Maximum capacity “Max”</i>	<i>Lower limit of the minimum load “Min”</i>	<i>Scale interval “d”</i>	<i>Number of scale intervals “n”</i>	<i>Verification scale interval, “e”</i>
Non-graduated machines				
400 g ≤ Max < 2 kg	10 e			5 g
2 kg ≤ Max	10 e			Max400
Graduated machines				
<i>Non-self indicating machines</i>				
400 g ≤ Max < 2 kg	10 d	5 g ≤ d ≤ 10 g	80 ≤ n < 400	5 g
2 kg ≤ Max < 4 t	10 d	10 g ≤ d ≤ 10 kg	100 ≤ n < 400	Max400
2 kg ≤ Max ≤ 10 t	10 d	5 g ≤ d ≤ 10 kg	400 ≤ n ≤ 1 000	d
4 t ≤ Max	10 d	20 kg ≤ d	200 ≤ n < 400	Max400
8 t ≤ Max	10 d	20 kg ≤ d	400 ≤ n ≤ 1 000	d

Status: This is the original version (as it was originally made).

<i>Maximum capacity "Max"</i>	<i>Lower limit of the minimum load "Min"</i>	<i>Scale interval "d"</i>	<i>Number of scale intervals "n"</i>	<i>Verification scale interval, "e"</i>
<i>Self and semi-self indicating machines</i>				
$400 \text{ g} \leq \text{Max} \leq 10 \text{ t}$	10 d	$2 \text{ g} \leq d \leq 10 \text{ kg}$	$60 \leq n \leq 1\,000$	d
$4 \text{ t} \leq \text{Max}$	10 d	$20 \text{ kg} \leq d$	$200 \leq n \leq 1\,000$	d