SCHEDULE 1

ESSENTIAL REQUIREMENTS

Rated Operating Conditions

15. The manufacturer shall specify the rated operating conditions for the instrument as follows—

(a) For the measurand—

the measuring range in terms of its maximum and minimum capacity.

(b) For the electrical supply influence quantities—

(i) in case of AC voltage supply, the nominal AC voltage supply, or the AC voltage limits;

(ii) in case of DC voltage supply, the nominal and minimum DC voltage supply, or the DC voltage limits.

(c) For the mechanical influence quantities—

where the instrument is used under special mechanical strain, e.g. instruments incorporated into vehicles, the manufacturer shall define the mechanical conditions of use.

(d) For other influence quantities (if applicable)—

(i) the rate(s) of operation;

(ii) the characteristics of the product(s) to be weighed.