
STATUTORY INSTRUMENTS

2006 No. 1268

The Measuring Instruments (Cold-water Meters) Regulations 2006

PART III

USE FOR TRADE OF COLD-WATER METERS

Manner of use

- 16.**—(1) Where a cold-water meter is marked with—
- (a) a temperature range, it shall not be used for trade in temperatures outside that range;
 - (b) a flowrate range, it shall not be used for trade at a flowrate outside that range.
- (2) Where a cold-water meter bears a mark which signifies the manner and purpose of use, it shall not be used for trade in a manner or for a purpose which does not accord with that marking.
- (3) A cold-water meter shall not be used for trade in circumstances—
- (a) in which it may be prevented from operating consistently or accurately; or
 - (b) which are likely prematurely to degrade its metrological characteristics.
- (4) A cold-water meter shall not be used for trade unless, when adjusted, the calibration of the instrument is set as close to zero as is practicable.