

SCHEDULE 1

Regulation 9(2)

INFORMATION TO BE MARKED ON REGULATED
NON-AUTOMATIC WEIGHING INSTRUMENTS

1. The number of the EU-type examination certificate, where appropriate.
2. The manufacturer's name, registered trade name or registered trade mark.
3. The accuracy class, enclosed in an oval or in two horizontal lines joined by two half circles.
4. Maximum capacity, in the form "Max".
5. Minimum capacity, in the form "Min".
6. Verification scale interval in the form "e =....."
7. Type, batch and serial number
8. When applicable the following:
 - (a) for instruments consisting of separate but associated units, the identification mark on each unit;
 - (b) scale interval if different from e, in the form "d =....";
 - (c) maximum additive tare effect, in the form "T = +.....";
 - (d) maximum subtractive tare effect if it is different from Max, in the form "T = -...";
 - (e) tare interval if it is different from d, in the form "d_T =";
 - (f) maximum safe load if it is different from "Max....", in the form "Lim....";
 - (g) the special temperature limits, in the form "...°C/...°C"; and
 - (h) ratio between load receptor and load.
9. The requirements of points 1.2 to 1.5 of the Annex III to the Directive must be complied with.