
STATUTORY INSTRUMENTS

2016 No. 1152

The Non-automatic Weighing Instruments Regulations 2016

PART 3

**CONFORMITY OF REGULATED NON-
AUTOMATIC WEIGHING INSTRUMENTS**

CHAPTER 1

ESTABLISHING COMPLIANCE WITH THE ESSENTIAL REQUIREMENTS

Conformity assessment procedures

36.—(1) The conformity of regulated non-automatic weighing instruments to the essential requirements may, subject to paragraph (2), be established by either of the following conformity assessment procedures as selected by the manufacturer—

(a) Module B as set out in point 1 of Annex II to the Directive followed by either—

(i) Module D as set out in point 2 of Annex II; or

(ii) Module F as set out in point 4 of Annex II; or

(b) Module G as set out in point 6 of Annex II.

(2) Module B is compulsory for instruments—

(a) which use electronic devices; and

(b) the load measuring device of which uses a spring to balance the load.

(3) Where an instrument is not submitted to Module B, either of the following modules must be applied—

(a) Module D1 as set out in point 3 of Annex II; or

(b) Module F1 as set out in point 5 of Annex II.

(4) A notified body must carry out the conformity assessment procedure selected by the manufacturer in accordance with the requirements of Schedule 2.

(5) The documents and correspondence relating to the conformity assessment procedures referred to in this regulation, and which are carried out in the United Kingdom, must be drawn up in English.

(6) In this regulation a reference to a module other than Module B includes the common provisions as set out in point 7 of Annex II to the Directive.