

ANNEX

DETERMINATION OF DICHLOROMETHANE AND 1,1,1-TRICHLOROETHANE

5. APPARATUS

- 5.1. Usual laboratory apparatus.
- 5.2. Gas chromatograph fitted with a thermal conductivity detector.
- 5.3. Transfer bottle, 50 to 100 ml (see sampling method 5.3)⁽¹⁾.
- 5.4. Pressure gas-syringe, 25 or 50 µl (see sampling method 5.4.2.2)⁽¹⁾.

Status: *This is the original version (as it was originally adopted).*

- (1) [OJ No L 383, 31. 12. 1980, p. 27.](#)