

## ANNEX III

## INSCRIPTIONS

**1. Instruments intended to be used for the applications listed in points (a) to (f) of Article 1(2)**

1.1. Those instruments shall bear visibly, legibly and indelibly the following inscriptions:

- (i) the number of the EU-type examination certificate, where appropriate;
- (ii) the manufacturer's name, registered trade name or registered trade mark;
- (iii) the accuracy class, enclosed in an oval or in two horizontal lines joined by two half circles;
- (iv) maximum capacity, in the form Max ...;
- (v) minimum capacity, in the form Min ...;
- (vi) verification scale interval, in the form  $e = \dots$ ;
- (vii) type, batch or serial number;  
and when applicable:
  - (viii) for instruments consisting of separate but associated units: identification mark on each unit;
  - (ix) scale interval if it is different from  $e$ , in the form  $d = \dots$ ;
  - (x) maximum additive tare effect, in the form  $T = + \dots$ ;
  - (xi) maximum subtractive tare effect if it is different from Max, in the form  $T = - \dots$ ;
  - (xii) tare interval if it is different from  $d$ , in the form  $d_T = \dots$ ;
  - (xiii) maximum safe load if it is different from Max, in the form Lim ...;
  - (xiv) the special temperature limits, in the form ... °C/... °C;
  - (xv) ratio between load receptor and load.