

ANNEX VIII

SPECIFIC REQUIREMENTS

CHAPTER IV

Discontinuous Totalisers

8. **Performance under influence factors and electromagnetic disturbances**
- 8.2. The critical change value due to a disturbance is one totalisation scale interval for any weight indication and any stored total.