

ANNEX X

MATERIAL MEASURES (MI-008)

CHAPTER II

Capacity serving measures

SPECIFIC REQUIREMENTS

4. **Shape**
- 4.1. Transfer measures shall be designed so that a change of contents equal to the MPE causes a change in level of at least 2 mm at the brim or filling mark.
- 4.2. Transfer measures shall be designed so that the complete discharge of the liquid being measured will not be impeded.