

Commission Regulation (EU) No 601/2012 of 21 June 2012 on the monitoring and reporting of greenhouse gas emissions pursuant to Directive 2003/87/EC of the European Parliament and of the Council (Text with EEA relevance) (repealed)

CHAPTER III

MONITORING OF EMISSIONS OF STATIONARY INSTALLATIONS

SECTION 3

Measurement-based methodology

Article 42

Measurement standards and laboratories

1 All measurements shall be carried out applying methods based on EN 14181 Stationary source emissions — Quality assurance of automated measuring systems, EN 15259 Air quality — Measurement of stationary source emissions — Requirements for measurement sections and sites and for the measurement objective, plan and report, and other corresponding EN standards.

Where such standards are not available, the methods shall be based on suitable ISO standards, standards published by the Commission or national standards. Where no applicable published standards exist, suitable draft standards, industry best practice guidelines or other scientifically proven methodologies shall be used, limiting sampling and measurement bias.

The operator shall consider all relevant aspects of the continuous measurement system, including the location of the equipment, calibration, measurement, quality assurance and quality control.

2 The operator shall ensure that laboratories carrying out measurements, calibrations and relevant equipment assessments for continuous emission measurement systems (CEMS) shall be accredited in accordance with EN ISO/IEC 17025 for the relevant analytical methods or calibration activities.

Where the laboratory does not have such accreditation, the operator shall ensure that equivalent requirements of Article 34(2) and (3) are met.

Changes to legislation:

There are currently no known outstanding effects for the Commission Regulation (EU) No 601/2012 (repealed), Article 42.