

SCHEDULE 4

Matters in respect of which a radiation protection adviser shall be consulted

3. The regular calibration of equipment provided for monitoring levels of ionising radiation and the regular checking that such equipment is serviceable and correctly used.