
STATUTORY INSTRUMENTS

1995 No. 735

**The Measuring Equipment (Capacity Measures
and Testing Equipment) Regulations 1995**

PART III

TESTING

Method of testing of capacity measures

- 9.** Subject to regulation 12 below, the inspector shall test every capacity measure—
- (a) by trial for use in the manner in which it is to be used as a measure, either for holding or for delivering its contents;
 - (b) by using—
 - (i) where practicable, water or an intended liquid of use, or
 - (ii) in the case of a dry capacity measure where the use of a liquid is impracticable, seeds, glass beads or polystyrene beads of an appropriate and regular size, proper precautions being taken to achieve reliable and consistent results; and
 - (c) subject in the case of dispensing measures for pharmaceutical purposes to regulation 11 below, either—
 - (i) volumetrically, by comparison with one or more local or working standards or by means of appropriate testing equipment which complies with the requirements of the Weights and Measures (Local and Working Standard Capacity Measures and Testing Equipment) Regulations 1990⁽¹⁾; or
 - (ii) gravimetrically, by determining the weight of water of known density using a suitable weighing machine.

Commencement Information

II Reg. 9 in force at 10.4.1995, see **reg. 1(1)**

⁽¹⁾ S.I.1990/2626 as amended by regulation 19 of these Regulations.

Status:

Point in time view as at 10/04/1995.

Changes to legislation:

There are currently no known outstanding effects for the The Measuring Equipment (Capacity Measures and Testing Equipment) Regulations 1995, Section 9.