

SCHEDULE 1

Regulation 1(2)

REVOCATIONS

Commencement Information**II** Sch. 1 in force at 10.4.1995, see **reg. 1(1)**

<i>Regulation revoked</i>	<i>Reference</i>	<i>Extent of revocation</i>
The Weights and Measures Regulations 1963	S.I. 1963/1710	Regulation 1(1)(b) and (c) and (2)(f), regulation 3, regulation 5(1)(b)(i) and (vii), regulation 6(2)(b) to (d), Parts III and IV and Schedule 1, Part II
The Weights and Measures (Amendment) Regulation 1970	S.I. 1970/1711	The whole Regulations so far as unrevoked
The Weights and Measures (Amendment) Regulations 1974	S.I. 1974/1326	The whole Regulations

SCHEDULE 2

Regulation 15

PRESCRIBED LIMITS OF ERROR FOR CAPACITY MEASURES OTHER THAN DISPENSING MEASURES FOR PHARMACEUTICAL PURPOSES

Commencement Information**I2** Sch. 2 in force at 10.4.1995, see **reg. 1(1)**

<i>Normal Capacity</i>		<i>Limits for error</i>	
<i>Imperial</i>	<i>Metric</i>	<i>Line measure, in excess and in deficiency</i>	<i>Brim measure, in excess only</i>
(1)	(2)	(3)	(4)
5, 4 or 3 gal	20 l	100 ml	200 ml

* for measuring intoxicating liquor only.

Notes:

- The limit of error in relation to any scale mark on a sub-divided or graduated measure shall be that specified for a measure of nominal capacity equivalent to the amount which the scale mark is intended to denote, or where the nominal capacity equivalent to the amount does not appear in this Schedule it shall be calculated by interpolation.
- On and after 1st October 1995—
 - the foregoing table shall have effect with the omission in column 1 of the references to 5, 4 or 3 gal, 2 gal, 1 gal, [^F1½ gal,] 1 qt, 8 fl oz, 6 fl oz, 1 gill, 4 fl oz, ½ gill, #, ⅓ gill, ¼, # gill and # gill; and
 - references to imperial measures shall be to their use for measuring beer and cider only.

Status: Point in time view as at 03/04/2001.

Changes to legislation: There are currently no known outstanding effects for the The Measuring Equipment (Capacity Measures and Testing Equipment) Regulations 1995. (See end of Document for details)

<i>Normal Capacity Imperial</i>	<i>Metric</i>	<i>Limits for error Line measure, in excess and in deficiency</i>	<i>Brim measure, in excess only</i>
(1)	(2)	(3)	(4)
2 gal, 16 pt	10 l	50 ml	100 ml
1 gal, 8 pt	5 l	25 ml	50 ml
½ gal, 4 pt	2.5 or 2 l	12 ml	24 ml
1 qt, 2 pt	1 l	10 ml	20 ml
1 pt	500 ml	5 ml	10 ml
½ pt	250 ml	4 ml	8 ml
8 fl oz or ⅓ pt	200 ml	4 ml	8 ml
6 fl oz	175 ml *	3.5 ml	7 ml
1 gill	150 ml *	3 ml	6 ml
4 fl oz	125 ml *	2.5 ml	5 ml
½ gill	[^{F2} 100 ml, 70 ml*]	2 ml	4 ml
#, ⅓ gill	50 ml	2 ml	4 ml
¼, # gill	25 ml, 35 ml *	2 ml	4 ml
# gill	20 ml	2 ml	4 ml
	10 ml	1 ml	2 ml
	4 ml	0.5 ml	1 ml
	2 ml	0.2 ml	0.4 ml
	1 ml	0.1 ml	0.2 ml

* for measuring intoxicating liquor only.

Notes:

1. The limit of error in relation to any scale mark on a sub-divided or graduated measure shall be that specified for a measure of nominal capacity equivalent to the amount which the scale mark is intended to denote, or where the nominal capacity equivalent to the amount does not appear in this Schedule it shall be calculated by interpolation.
2. On and after 1st October 1995—
 - (a) the foregoing table shall have effect with the omission in column 1 of the references to 5, 4 or 3 gal, 2 gal, 1 gal, [^{F1}½ gal,] 1 qt, 8 fl oz, 6 fl oz, 1 gill, 4 fl oz, ½ gill, #, ⅓ gill, ¼, # gill and # gill; and
 - (b) references to imperial measures shall be to their use for measuring beer and cider only.

F1 Words in Sch. 2 Note 2 inserted (3.4.2001) by [The Measuring Equipment \(Capacity Measures\) \(Amendment\) Regulations 2001 \(S.I. 2001/599\)](#), regs. 1, **4(e)**

F2 Words in Sch. 2 Table substituted (3.4.2001) by [The Measuring Equipment \(Capacity Measures\) \(Amendment\) Regulations 2001 \(S.I. 2001/599\)](#), regs. 1, **4(d)**

Status:

Point in time view as at 03/04/2001.

Changes to legislation:

There are currently no known outstanding effects for the The Measuring Equipment (Capacity Measures and Testing Equipment) Regulations 1995.