SCHEDULE 2

METHODS OF ANALYSIS

PART I

25b.

EXTRACTION OF WATER-SOLUBLE TRACE ELEMENTS

5.2 APPARATUS

5.2. pH-meter.

(Note) Where the boron content of the extract is to be determined, do not use borosilicate glassware. Teflon or silica is preferable for this extraction. Rinse the glassware thoroughly if it has been washed in detergents containing borates.