

SCHEDULE 2

ANNEX I TO COUNCIL DIRECTIVE [90/384/EEC](#) of 20TH JUNE 1990 ON THE HARMONISATION OF THE LAWS OF MEMBER STATES RELATING TO NON-AUTOMATIC WEIGHING INSTRUMENTS METROLOGICAL REQUIREMENTS

Instruments with multiple weighing ranges

3.2. Multiple weighing ranges are permitted, provided they are clearly indicated on the instrument. Each individual weighing range is classified according to 3.1. If the weighing ranges fall into different accuracy classes the instrument shall comply with the severest of the requirements that apply for the accuracy classes in which the weighing ranges fall.