

SCHEDULE 2

ANNEX I TO COUNCIL DIRECTIVE [90/384/EEC](#) of 20TH JUNE 1990 ON THE HARMONISATION OF THE LAWS OF MEMBER STATES RELATING TO NON-AUTOMATIC WEIGHING INSTRUMENTS

METROLOGICAL REQUIREMENTS

Multi-interval instruments

3.3.1. Instruments with one weighing range may have several partial weighing ranges (multi-interval instruments).

Multi-interval instruments shall not be equipped with an auxiliary indicating device.