

SCHEDULE 2

ANNEX I TO COUNCIL DIRECTIVE [90/384/EEC](#) of 20TH JUNE 1990 ON THE HARMONISATION OF THE LAWS OF MEMBER STATES RELATING TO NON-AUTOMATIC WEIGHING INSTRUMENTS METROLOGICAL REQUIREMENTS

Accuracy

4.2. The maximum permissible errors in service are twice the maximum permissible errors fixed in section 4.1.