

SCHEDULE 3

(Requirements in respect of Use and Manner of Use)

Zero-tracking device (Extract from Clause 3.3.2 of OIML R51)

A zero-tracking device shall operate only when:

- the indication is at zero, or at a negative net value equivalent to gross zero, and
- the stability criteria are fulfilled, and
- the corrections are not more than 0.5 e/second.