

SCHEDULE 6

IN SERVICE REQUIREMENTS FOR CERTAIN REGULATED MEASURING INSTRUMENTS IN GREAT BRITAIN

PART 9

BELTWEIGHERS

Requirements for use for trade

34. No person shall use for trade a beltweigher unless—

- (a) it is compliant with the essential requirements other than the provisions relating to maximum permissible errors;
- (b) the requirements of paragraphs 35 and 36 are complied with;
- (c) it has been erected and installed in accordance with the requirements of paragraph 37; and
- (d) in the case of a beltweigher which is stated to be of an accuracy class set out in column 1 of the Table set out in this paragraph it is within the maximum permissible error for that accuracy class as set out in column 2 of that Table.

<i>Column (1)</i>	<i>Column 2</i>
Accuracy Class	Maximum permissible error for totalised load
0.5	± 0.5%
1	± 1.0%
2	± 2.0%