
STATUTORY INSTRUMENTS

2021 No. 905

The Medical Devices (Northern
Ireland Protocol) Regulations 2021

PART 8

Amendment of the Medical Devices Regulations 2002

Amendment of regulation 24A (UK(NI) indication: active implantable medical devices)

36. In regulation 24A, after paragraph (3), insert—

“(3A) The UK(NI) indication may be less than 5mm high provided that it is the same height as the CE marking that it accompanies.”

Commencement Information

11 Reg. 36 in force at 27.7.2021, see [reg. 1\(2\)](#)

Changes to legislation:

There are currently no known outstanding effects for the The Medical Devices (Northern Ireland Protocol) Regulations 2021, Section 36.